

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

Supervisor/Operator Performing the Verification Procedure:

A Name Robert J. Ely ID# 2681 Date 02-14-2012
Agency Craig Police Department Phone # (907) 826-3330
Instrument Location 506 2nd Street, Craig, Alaska 99921

B Alco S/N 75509 Target Value .079 High Pressure 400 PSI
Alco Test Value Average .081 .081
1st Alco 2nd Alco

Signature Robert J. Ely

(OVER)

(Do Not write in the area below)

EMB
2/22/11

I, Nita J. Bolz, after being first duly sworn, depose and state as follows;

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons Responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012 4/5/12

Nikki Roth (Notary Seal Stamp)
~~Daniel L. Nicholas~~ Nikki Roth
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130195

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

FEBRUARY 14, 2012

OPERATOR'S NAME:

ELY, RJ

OPERATOR'S NUMBER: 2681

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VER OF CAL

O.L. #: 123456789

DEPT/AGENCY: CRG1

CASE/REPORT: 12-XXXXX

TEST TYPE: V

ALCO TARGET VALUE: .079

ALCO S/N: 75509

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.22 in 08:07
ALCO TARGET .079 08:08
BLANK TEST .000 08:08
INTERNAL STANDARD VERIFIED 08:08
ALCO TV 30.22 in .081 08:09
BLANK TEST .000 08:09
SUBJECT SAMPLE .000 08:10
BLANK TEST .000 08:11
ALCO TV 30.22 in .081 08:11
BLANK TEST .000 08:12

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130195

FEBRUARY 14, 2012

TIME 08:05

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/14/12

HEATERS: OKAY

SAMPLE: OKAY

PUMP: OKAY

BAROMETER: 30.22 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

ALCO S/N: 0123456789; KK=70AB0CDEF6
4123456789; JJJ=70AB0CDEF6
9876543210; LLL=70AB0CDEF6